



**LEMBAGA GETAH MALAYSIA**  
**MALAYSIAN RUBBER BOARD**  
**RUBBER RESEARCH INSTITUTE OF MALAYSIA**

**Unit Pencirian Bahan**  
 260 Jalan Ampang, 50450 Kuala Lumpur.  
 Peti Surat 10150, 50908 Kuala Lumpur, Malaysia.  
 Tel: (6)03-92062000, 92063681 Fax: (6)03-42519213  
 Email: kartinia@lgm.gov.my




MS,ISO/IEC 17025  
 TESTING  
 SMM NO.069

Our report no. : **LGM/BTK/UPB/5.10/CP/0902/0032**  
 Your reference no. :-

**February 13, 2009**

**Sanctuary Health Sdn. Bhd.**  
**464, Jalan Kuala Kangsar**  
**30010 Ipoh**  
**Perak DR, Malaysia**



**Fax : 605-2921739**

**Attn: Chang Chaur Long**



**ANALYTICAL TEST RESULTS**

We report the following results:

Your letter dated 23/01/2009  
 Date of sample received 03/02/2009  
 Date of testing 12/02/2009  
 Test method In House Method Following ASTM D6124-01  
 Type of Sample Sanctuary Dam  
 Lot no 181205  
 Size 0.18mm x 5" x 5"  
 Quantity Tested 4 pcs

No.	Lab. Reference No.	Sample Reference	Residual Powder Content (mg/glove)
1.	0032	Colour Green	0.10

\* Values obtained are for whole piece of sample

**"BERKHIDMAT UNTUK NEGARA"**

**Kartini Alias**  
**Research Officer**  
**Material Characterization Unit**  
**Malaysian Rubber Board**  
**For Director General**

The uncertainties are for a confidence probability of no less than 95%

This report is issued in accordance with the conditions of accreditation granted by SMM which has assessed the capability of the laboratory and its traceability to recognised national standards and to the units of measurement realised at the corresponding national standards laboratory. Copyright of this report is owned by the issuing laboratory and may not be reproduced other than in full except with prior written approval of the Head of the issuing laboratory.